

## ADDENDUM No.1

This is in reference to Tender No. MOALF&C/SDCDAR/NARIGP/002/2020-2021 and Tender No. **MOALF&C/SDCDAR/NARIGP/003/2020-21 for Supply, Delivery, Installation, Training commissioning of Lab Equipment and Supply, Delivery, Installation, Training commissioning and Maintenance of Unified Digital Platform** which were published on Standard Newspapers of 30<sup>th</sup> March, 2021.

In response to the clarifications sought by the participating firms in accordance with clause 8 of the ITB of the issued bidding document, the employer wishes to clarify on the issues raised as follows

<b>TENDER No. MOALF&amp;C/SDCDAR/NARIGP/002/2020-21 : Supply delivery installation and commissioning of lab equipment</b>			
<b>No</b>	<b>Issues raised</b>	<b>Clarifications sought</b>	<b>clarifications</b>
1	List of goods to be supplied on page 62 differ with quantities on page 64-68	Clarify the correct quantities	a)The correct quantity is as per the quantities indicated in the technical specifications on page 64-68 <b>BUT</b> b)Quantities for color sorter accessories are three as indicated under clarification No. 3
2	Project sites are three as indicated SCC CLAUSE 28.3 ON PAGE 3-103	Provide list of goods & delivery schedule for Sagana and Meru	All items will be delivered and received at the New KPCU Dandora.
3	a) Quantities of color sorter and color sorter accessories differ  b) Installation roof height	Clarify if You require three color sorters but only two accessories ie feeding bucket elevator, On-Floor STUDY metallic inlet hoper, screw compressor, dust extraction and collection system  Confirm the installation roof height to assist on sizing the bucket elevator	The correct quantity of each accessory listed for color sorters is <b>three</b> and not two.  The maximum stand height of seven (7) metres and necessary accessories to be compatible with seven (7m) height
4	Under technical specifications has column for unit price in US\$, KES and total price in KES.	Confirm if this is in order as the same price pricing shall be depicted in underprice schedule	This is ok. Quote the unit price and the total price in <b>Kes</b>
4	Laboratory tables	Assist on information on how / view the sample of the lab tables	The Laboratory tables can be viewed at New KPCU Dandora The specifications are (4ft x8ft top measurements, 3.2ft height with shelves.

**TENDER No. MOALF&C/SDCDAR/NARIGP/003/2020-21: Supply delivery Installation, training commissioning and Maintenance of Unified Digital Platform**

1	Bid Security & form of Bid	Clarify if one has to attach both Bid security (bank guarantee) and form of bid security.	No. Provide a bid security in the form of bank guarantee
2	In addition to the original bid the No. of copies is one (1) On page 92 –two copies of financial & technical proposal	Clarify the correct No. of copies required	In addition to the original, provide <b>one (1)</b> copy of technical and financial proposal
3	On page 37, there is a requirement of five past contracts and page 92 the requirement is at least three past contracts	Clarify the correct number of past contracts	The No. of contracts required is three (3) past contracts of similar scope.

**N/B:**

1. Other terms and conditions of the two tenders remains the same.
2. This addendum is part of the bidding documents
3. The addendum can also be downloaded from Kilimo website: [www.kilimo.go.ke](http://www.kilimo.go.ke)